

ImmunoFluoroMount™

Cat. No. GTX30928

Applications

IHC

References (5)

Package

30 ml

PRODUCT

Refractive Index: 1.400 ± 0.002

Summary

(These numbers apply to these mounting mediums in solution. Refractive indexes change when the water solvent evaporates and mounting media dries on slides. We do not have the means to measure the refractive indexes of dry mounting mediums; however we expect the numbers to go higher when dried. The refractive index of water is 1.3330)

Applications

Application Note

Apply 2-3 drops of mounting medium and apply coverslip carefully without getting air bubbles.

Properties

Form

Liquid

Storage

Store at 2-8°C, Protect from light, DO NOT FREEZE.

For laboratory research use only. Not for any clinical, therapeutic, or diagnostic use in humans or animals. Not for animal or human consumption.

Note

Purchasers shall not, and agree not to enable third parties to, analyze, copy, reverse engineer or otherwise attempt to determine the structure or sequence of the product.



For full product information, images and publications, please visit our [website](#).

Date 2026 / 02 / 01 Page 1 of 1