

Androgen Receptor antibody [DHTR/882]

Cat. No. GTX34414

Host	Mouse	Package
Clonality	Monoclonal	100 µg
Isotype	IgG1	
Applications	IHC-P	
Reactivity	Human	

Applications

Application Note

*Optimal dilutions/concentrations should be determined by the researcher.

Suggested dilution	Recommended dilution
IHC-P	1-2µg/ml for 30 minutes at RT

Note : Staining of formalin-fixed tissues requires heating tissue sections in 10mM Tris with 1mM EDTA, pH 9.0, for 45 min at 95°C followed by cooling at RT for 20 minutes.

Not tested in other applications.

Product Note This antibody reacts with full length, and the newly described A form of the receptor. It does not cross react with estrogen, progesterone, or glucocorticoid receptors.
We do not recommend use of this product for Mouse samples.

Properties


Form	Liquid
Buffer	PBS, 0.05% BSA
Preservative	0.05% Sodium azide
Storage	Store as concentrated solution. Centrifuge briefly prior to opening vial. For short-term storage (1-2 weeks), store at 4°C. For long-term storage, aliquot and store at -20°C or below. Avoid multiple freeze-thaw cycles.
Concentration	0.2 mg/ml (Please refer to the vial label for the specific concentration.)
Immunogen	Human DHTR recombinant protein fragment; aa151-350 (Exact sequence is proprietary)
Purification	Protein A/G purified
Conjugation	Unconjugated
Note	For laboratory research use only. Not for any clinical, therapeutic, or diagnostic use in humans or animals. Not for animal or human consumption. Purchasers shall not, and agree not to enable third parties to, analyze, copy, reverse engineer or otherwise attempt to determine the structure or sequence of the product.



For full product information, images and publications, please visit our [website](#).

Date 2026 / 01 / 10 Page 1 of 2

DATA IMAGES

**GTx34414 IHC-P Image**

IHC-P analysis of human prostate carcinoma tissue using GTx34414 Androgen Receptor antibody [DHTR/882].



For full product information, images and publications, please visit our [website](#).

Date 2026 / 01 / 10 Page 2 of 2