SARS-CoV-2 (COVID-19) Spike-ACE2 Binding / Neutralization Assay Kit

Cat. No. GTX536401

Applications	Neutralizing/Inhibition
Reactivity	SARS Coronavirus 2

References (1)
Package
96 test

PRODUCT

Summary

The GeneTex SARS-CoV-2 (COVID-19) Spike-ACE2 Binding / Neutralization Assay Kit is an in vitro assay for qualitative SARS-CoV-2 neutralization reagent screening. The kit includes the recombinant RBD proteins of the wildtype (WT) strain and five major WHO-designated variants of concern (i.e., Alpha, Beta, Gamma, Delta, and Kappa) for strain-specific neutralization analysis.

Applications

Application Note

The suitable sample types include but not limit to purified antibodies and serum. The suggested sample loading amount is 50µl per well. Please see the protocol for detailed procedure.

Properties

Storage	The entire kit can be stored at 2-8ºC for up to 6 months.
Conjugation	Horseradish peroxidase(HRP)
Note	For laboratory research use only. Not for any clinical, therapeutic, or diagnostic use in humans or animals. Not for animal or human consumption. Purchasers shall not, and agree not to enable third parties to, analyze, copy, reverse engineer or otherwise attempt to determine the structure or sequence of the product.

DATA IMAGES



GTX536401 Neutralizing/Inhibition Image

Inhibition analysis of SARS-CoV-2 (COVID-19) spike neutralizing control antibody using SARS-CoV-2 (COVID-19) Spike-ACE2 Binding / Neutralization Assay Kit (GTX536401).

For full product information, images and publications, please visit our website.

GeneTex, Inc. (North America)
1-877-436-3839 (Toll-free) 1-949-309-2888 sales@genetex.com

GeneTex International Corporation (Global)
886-3-6208988 886-3-6208989 infoasia@genetex.com

© 2018 GeneTex Inc. All rights reserved.



	RBD WT	RBD Alpha	RBD Beta	RBD Gamma	RBD Delta	RBD Kappa
Average (%)	99.7	94.1	108.3	99.2	104.7	99.9
Range (%)	87-115	79-110	100-116	93-107	91-120	93-112

GTX536401 Image

The recovery of SARS-CoV-2 (COVID-19) Spike-ACE2 Binding / Neutralization Assay Kit (GTX536401) was determined by spiking a spike neutralizing control antibody in human serum.

GTX536401 Image

GeneTex, Inc. (North America)
1-877-436-3839 (Toll-free) 1-949-309-2888 sales@genetex.com

GeneTex International Corporation (Global)
886-3-6208988 886-3-6208989 infoasia@genetex.com

© 2018 GeneTex Inc. All rights reserved.